


AD No. 2826B

FILE COPY

ASTORIA

WOODS HOLE OCEANOGRAPHIC INSTITUTION

Research



WOODS HOLE, MASSACHUSETTS

THIS REPORT HAS BEEN DELIMITED
AND CLEARED FOR PUBLIC RELEASE
UNDER DOD DIRECTIVE 5200.20 AND
NO RESTRICTIONS ARE IMPOSED UPON
ITS USE AND DISCLOSURE.

DISTRIBUTION STATEMENT A

APPROVED FOR PUBLIC RELEASE;
DISTRIBUTION UNLIMITED.

WOODS HOLE OCEANOGRAPHIC INSTITUTION

Woods Hole, Massachusetts

In citing this manuscript in a bibliography,
the reference should be followed by the
phrase: UNPUBLISHED MANUSCRIPT

Reference No. 54-17

A Suggestion for the Correction of
Salinity Data Obtained with
the S-T-D Instrument


by

Arthur R. Miller

Technical Report
Submitted to Geophysics Branch, Office of Naval Research
Under Contract N6onr-27701 (NR-083-004)

March 1954

APPROVED FOR DISTRIBUTION


Director

A SUGGESTION FOR THE CORRECTION OF SALINITY DATA
OBTAINED WITH THE S-T-D INSTRUMENT

Anyone who has compared titrated salinity samples with the results of automatic devices which measure salinity or conductivity of sea water electronically, has been aware of inadequacies in the application of a constant correction to the data. The purpose of this article is to demonstrate a type of correction which varies according to temperature and salinity. This correction is easily applied to data obtained with the Salinity-Temperature-Depth instrument (Jacobson, 1948). By this means, reconciliation of STD data with corresponding salinity samples has been accomplished with marked success for a wide range of conditions and stringency in the region about Delaware Bay.

Certain difficulties in the reconciliation of STD data with independent observations may be ascribed to maladjustment or misalignment of the mechanical linkage between the temperature circuit and the salinity computing circuit. This fault may not be readily apparent unless fairly wide ranges of salinity and temperature are encountered. Errors due to misalignment are made manifest by small discrepancies between check samples and STD record when salinity is low and/or temperature is high, expanding to large discrepancies when salinity is high and/or temperature is low.

It is readily seen that this phenomenon could seriously affect interpretations of STD data. For instance, in the case of a subsurface lowering, with a check sample obtained only at the surface, the ordinary reconciliation might not take into account the possibility of an increasing anomaly as the instrument was lowered through colder and more saline water. Lack of accuracy of the data offsets the advantage and convenience of the STD for rapid surveying if discrepancies are large. It is extremely important that, once the instrument is calibrated, adjustments at sea be kept at a minimum and then only with great care. In the light of the following considerations and formulae, the STD instrument and its computing circuit has proven most accurate.

CHARACTERISTIC ERROR OF THE STD

The first consideration will be to show the relation which may appear between a titrated salinity sample and the recorded salinity as computed by the instrument. The empirical formula for the salinity computing circuit is

$$S = \frac{100,000}{25.661 + 0.73720T} - 348.87 \quad C^{1.0916} \quad (1)$$

where S is salinity in parts per thousand, T the temperature in degrees Fahrenheit, C is specific conductance in mhos per centimeter cube (Jacobson, 1948). The formula is an approximation designed to give the least amount of error in ranges of salinity from 30 ‰ to 35 ‰.

If sea water of a certain salinity and temperature is being measured by the instrument, conductivity of the sample must also be a certain value. Any error in the computed salinity, aside from some constant calibration error, will depend upon an error in temperature, assuming no change in conductivity. Thus,

$$\frac{dS}{dT} = - \frac{73720 C^{1.0916}}{(.73720T + 25.661)^2}$$

or, approx.

$$\Delta S = - \frac{73720 C^{1.0916} \Delta T}{(.73720T + 25.661)^2} \quad (2)$$

The error in salinity, ΔS is proportional to the temperature error, ΔT .

Substituting for $C^{1.0916}$ in (2)

$$\Delta S = - \frac{73720 S \Delta T}{2,336,373.75 + 60520.66T - 189.60T^2}$$

which can be simplified to approximately

$$\Delta S = \frac{389 S \Delta T}{(T - 354)(T + 35)} \quad (3)$$

The error in salinity is proportional to the true salinity, S, and is inversely related to temperature, T.

The salinity error can be serious if the computing temperature differs from the real temperature by only a small amount. The following table shows some calculations for ΔS for several salinities and temperatures based on an instrumental error of 1°F.

	35 ‰	30 ‰	25 ‰	Salinity
80°F.	0.43	0.37	0.31	
50°	0.53	0.45	0.38	
30°	0.65	0.55	0.46	

If the accuracy of the instrument is assumed to be comparable to that of titrated samples, the table above has real significance. To be comparable to titrated values, temperature must not be in error by more than 0.1°F. (the tabulated ΔS divided by 10). Even if the recorded temperature is reasonably accurate, it can be linked erroneously with the computing circuit; consequently, ΔT can be a large value without the observer being aware of the difficulty. Consequently, the greater the ΔT , the greater, also, will be the apparent randomness of the salinity error, as shown by the table and equation (3)*.

Since equation (3) requires simultaneous knowledge of the salinity error and temperature error to determine the true salinity of an observation, it is not adequate, except as a check, for reconciling STD data. The following discussion will attempt to develop an empirical method for the correction of STD data.

EMPIRICAL FORMULA FOR CORRECTING STD DATA

In the three coordinate system let $x = S_r$, the recorded salinity resulting from instrumental computation, with the condition that at $x = 0$,

$$S_r = 0 \text{ and } S = 0.$$

The difference, $S_r - S = \Delta S$, will be taken vertically along the y axis (as shown in Fig. 1). For a given temperature, T, the relation, $\frac{\Delta S}{S}$, will be constant for a constant temperature error, ΔT , in the computing circuit. Also, the ratio

* The accuracy of the STD is claimed to be within 0.3 ‰. The successful use of this method of correction in over a thousand stations suggests that the accuracy of the STD is much better than above. Indeed, it is possible that such latitude in the probable salinity error is due to inherent sensitivity in the calibration of the instrument and inadequate procedures for maintenance at sea.

$$\frac{\Delta S}{S + \Delta S} = \frac{\Delta S}{S_r} = \tan \theta = k$$

Consequently,

$$\Delta S_n = k S_{r_n}$$

Or

$$\Delta S_n = \frac{S_{r_n}}{S_{r_o}} \cdot k \cdot S_{r_o}$$

If S_{r_o} is used as a reference, for a set of data with a constant temperature error,

$$\Delta S_n = A \frac{S_{r_n}}{S_{r_o}}$$

for a given temperature, T_{r_o} .

On the z axis let $z = T_{r_o} - T_{r_n}$ where T_{r_o} is any temperature recorded by the STD, suitable for reference. The point $z = 0$, $y = \Delta S$, $x = S_{r_o}$ applies to a known error of a recorded salinity at a particular temperature. If we assume that the temperature circuit is in good working order, ΔT is a constant error in the computing circuit and $T_r \approx T$. From equation (3), ΔS is nearly inversely proportional to T_r ; the recorded temperature, or, T , in the ranges 30° to 80° F.

Therefore, as temperature decreases from T_{r_o} , ΔS will increase. ΔS at $T_{r_o} - T_{r_n}$ will differ from ΔS at T_{r_o} , C , by an amount, C' and

$$C' = (T_{r_o} - T_{r_n}) \tan \gamma \quad (\text{Figure 1})$$

Let

$$C = \Delta S - C' = A \frac{S_{r_n}}{S_{r_o}}$$




FIG. 1

Without formal proof

$$\tan \nu = \ell S_r \tan \theta$$

where ℓ is a constant and $\tan \theta = \frac{C}{S_{r_0}}$ at T_{r_0} .

By substitution, then

$$\frac{C'}{T_{r_0} - T_{r_n}} = \ell C$$

Or

$$C' = \ell C (T_{r_0} - T_{r_n})$$

and C' is proportional to C for any constant temperature.

Since

$$C = A \frac{S_{r_n}}{S_{r_0}}$$

Then

$$C' = \ell A \frac{S_{r_n}}{S_{r_0}} (T_{r_0} - T_{r_n})$$

Combining,

$$C + C' = \Delta S$$

$$\Delta S = \frac{S_{r_n}}{S_{r_0}} \left[A + B (T_{r_0} - T_{r_n}) \right] \quad (4)$$

where $B = \ell A$.

It is readily seen that as long as a reference S_r is maintained, the formula, (4), can be transposed to the proven formula, $C = \Delta S$, by substituting T_{r_n} for T_{r_0} .

For a set of STD data in which there are a number of calibration samples, formula (4) becomes

$$\Delta S = \frac{S_r}{\bar{S}_r} \bar{\Delta S} + B (\bar{T}_r - T_r)$$

Where $\bar{S}_r = S_{r_0}$, the average S_r for which there are samples, $\bar{T}_r = T_{r_0}$, the average T_r of the same samples, $\bar{\Delta S} = A$, the average deviation of S_r from S in the set of data. Once B is known, ΔS can be solved for any pair of values, S_r and T_r .

B is solved for in the equation

$$B = \sum \left(\frac{\frac{\Delta S_n \bar{S}_r}{S_{r_n}} - \bar{\Delta S}}{\bar{T}_r - T_{r_n}} \right) / n$$

Ideally, there should be no difficulty solving for B , but the sensitivity of the computing circuit to temperature is such that the accuracy needed to check the STD data is scarcely attainable at sea. For the most part, individual calculations of B are reasonably constant, except for $T_{r_0} - T_{r_n} < 1^\circ$, or, obviously inadequate sampling. The latter cases should not be included in the averaging of B . With the completion of the equation, it is a simple matter to construct a nomogram (Fig. 2) for the range of values in the set of data and to apply the indicated corrections.

The validity of the calculations is dependent on the accuracy, number and range of the checking samples. If a hidden constant calibration error is included in the calculations, the accuracy of the formula is affected percentage-wise. It is a matter of judgment whether to apply the formula or not, depending on the apparent randomness of the error.

REFERENCE

- Jacobson, A. W., 1948: An instrument for recording continuously the salinity, temperature and depth of sea water. Trans. Amer. Inst. Elect. Engrs., 67: 714-722.




FIG. 2

DISTRIBUTION LIST FOR UNCLASSIFIED TECHNICAL REPORTS
Contract N6onr-27701 (NR-C83-004)

page 1a

<u>Copies</u>	<u>Addressee</u>
1	Commanding Officer Air Force Cambridge Research Center 230 Albany Street Cambridge 39, Massachusetts Attn: CRHSL
1	Director Air University Library Maxwell AF Base, Alabama Attn: CR-5110
1	Allan Hancock Foundation University of Southern California Los Angeles 7, California
5	Armed Services Technical Information Center Documents Service Center Knott Building Dayton 2, Ohio
2	Assistant Naval Attache for Research American Embassy Navy Number 100 Fleet Post Office New York, New York
1	Assistant Secretary of Defense (R&D) Pentagon Building Washington 25, D. C. Attn: Committee on Geophysics and Geography
1	Dr. William H. Sutcliffe, Director Bermuda Biological Station for Research St. George's, Bermuda
1	Head Department of Oceanography Brown University Providence, Rhode Island
1	Librarian California Fisheries Laboratory Terminal Island Station San Pedro, California
1	Director Chesapeake Bay Institute Box 426A R. F. D. #2 Annapolis, Maryland

DISTRIBUTION LIST

page 2a

2 Chief, Bureau of Ships
Department of the Navy
Washington 25, D. C.
Attn: Code 847

1 Chief, Bureau of Yards and Docks
Department of the Navy
Washington 25, D. C.

1 Chief of Naval Operations (Op-533D)
Department of the Navy
Washington 25, D. C.

3 Chief of Naval Research
Department of the Navy
Washington 25, D. C.
Attn: Code 416 (2)
Code 466 (1)

1 Department of Conservation
Cornell University
Ithaca, New York
Attn: Dr. J. C. Ayers

1 Commanding General
Research and Development Division
Department of the Air Force
Washington 25, D. C.

1 Commanding General
Research and Development Division
Department of the Army
Washington 25, D. C.

1 The Oceanographic Institute
Florida State University
Tallahassee, Florida

1 Director
Lamont Geological Observatory
Torrey Cliff
Palisades, New York

1 Director
Narragansett Marine Laboratory
University of Rhode Island
Kingston, Rhode Island

1 National Research Council
2101 Constitution Avenue
Washington 25, D. C.
Attn: Committee on Undersea
Warfare

DISTRIBUTION LIST

page 3a

1	Commanding Officer Naval Ordnance Laboratory White Oak Silver Spring 19, Maryland
6	Director Naval Research Laboratory Washington 25, D. C. Attn: Technical Information Officer
1	Office of Naval Research Branch Office 1030 East Green Street Pasadena 1, California
1	Office of Naval Research Branch Office 1000 Geary Street San Francisco 9, California
1	Office of Naval Research Branch Office Tenth Floor, John Crerar Library Bldg. 86 East Randolph Street Chicago 11, Illinois
1	Office of Naval Research Branch Office 150 Causeway Street Boston 14, Massachusetts
1	Office of Naval Research Branch Office 346 Broadway New York 13, New York
2	Officer-in-Charge Office of Naval Research London Branch Office Navy Number 100 Fleet Post Office New York, New York
1	Office of Technical Services Department of Commerce Washington 25, D. C.
1	Dr. Willard J. Pierson New York University New York 53, New York
1	Department of Zoology Rutgers University New Brunswick, New Jersey Attn: Dr. H. H. Haskin
2	Director Scripps Institution of Oceanography La Jolla, California

L.

J

DISTRIBUTION LIST

page 4a

1	Head Department of Oceanography Texas A & M College Station, Texas
1	Department of Engineering University of California Berkeley, California
1	Director Hawaii Marine Laboratory University of Hawaii Honolulu, T. H.
1	Director Marine Laboratory University of Miami Coral Gables 34, Florida
1	Head Department of Oceanography University of Washington Seattle 5, Washington
1	U. S. Army Beach Erosion Board 5201 Little Falls Road, N. W. Washington 16, D. C.
1	Director U. S. Coast and Geodetic Survey Department of Commerce Washington 25, D. C.
1	Commandant (OFU) U. S. Coast Guard Washington 25, D. C.
1	U. S. Fish and Wildlife Service 450 B Jordan Hall Stanford University Stanford, California
1	U. S. Fish and Wildlife Service Fort Crockett Galveston, Texas
1	U. S. Fish and Wildlife Service P. O. Box 3830 Honolulu, T. H.
1	U. S. Fish and Wildlife Service Woods Hole Massachusetts

L

DISTRIBUTION LIST

page 5a

2 Director
U. S. Fish and Wildlife Service
Department of the Interior
Washington 25, D. C.
Attn: Dr. L. A. Walford

1 Project Arowa
U. S. Naval Air Station, Bldg. R-48
Norfolk, Virginia

1 Department of Aerology
U. S. Naval Post Graduate School
Monterey, California

2 Director
U. S. Navy Electronics Laboratory
San Diego 52, California
Attn: Code 550
Code 552

8 Hydrographer
U. S. Navy Hydrographic Office
Washington 25, D. C.
Attn: Division of Oceanography

1 Bingham Oceanographic Foundation
Yale University
New Haven, Connecticut